
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002268
9:58 15/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:52
PRESSURIZING 1:58
STERILIZING 2:51
136.0 C 220kPa 2:51
136.4 C 226kPa 3:21
136.2 C 224kPa 3:51
136.4 C 226kPa 4:21
136.4 C 225kPa 4:51
136.1 C 221kPa 5:21
136.4 C 226kPa 5:51
136.4 C 224kPa 6:21
136.4 C 224kPa 6:21
VENTING 6:21
AIR DRYING 7:12
CYCLE COMPLETE 37:12

Digital Signature #
8BCAE25ADEA04745
