
Midmark M11 - v1.0.5
Total Cycles: 2511

Name: RITTER2
Sterilizer ID: V2313228
06/14/2023 07:18 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.6 F 0.0
2:00 75.8 F 0.0
4:00 78.4 F 0.0
6:00 98.8 F 0.0
8:00 151.1 F 0.4
10:00 164.1 F 3.2
12:00 187.9 F 6.6
14:00 205.2 F 10.1
16:00 215.4 F 10.2
18:00 223.9 F 8.1
20:00 234.9 F 12.5
22:00 234.0 F 9.3
24:00 238.5 F 10.2
26:00 244.6 F 13.2
28:00 255.8 F 19.1
30:00 265.9 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.2
1:00 272.5 F 29.9
2:00 271.7 F 29.1
3:00 272.0 F 29.9
4:00 272.3 F 29.6

Min 271.3 F 29.0
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.0 F 0.0
5:00 215.3 F 0.0
10:00 213.2 F 0.0
15:00 227.5 F 0.0
20:00 229.1 F 0.0
25:00 229.5 F 0.0
30:00 230.4 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 31:45
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:09:42
CYCLE COMPLETE

06/14/2023 08:27 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____