
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001663
16:16 11/12/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:08
PRESSURIZING 2:14
STERILIZING 3:13
136.4 C 222kPa 3:13
136.3 C 225kPa 3:43
136.4 C 225kPa 4:13
136.4 C 226kPa 4:43
136.4 C 225kPa 5:13
136.4 C 227kPa 5:43
136.3 C 224kPa 6:13
136.3 C 225kPa 6:43
136.3 C 225kPa 6:43
VENTING 6:43
AIR DRYING 7:35
CYCLE COMPLETE 26:09

Digital Signature #
AC34DE4A266DCA78
