
Midmark M11 - v1.0.5
Total Cycles: 2181

Name: RITTER2
Sterilizer ID: V2313228
03/27/2023 03:51 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 162.3 F 0.0
2:00 167.2 F 3.3
4:00 189.5 F 7.4
6:00 200.6 F 12.6
8:00 217.6 F 6.8
10:00 233.9 F 12.2
12:00 234.6 F 9.4
14:00 237.6 F 9.4
16:00 248.2 F 14.8
18:00 260.7 F 21.8
20:00 271.8 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.1
1:00 272.6 F 29.8
2:00 271.6 F 29.0
3:00 272.6 F 29.9
4:00 272.2 F 29.4

Min 271.4 F 28.9
Max 272.8 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.1 F 0.0
5:00 225.2 F 0.0
10:00 238.4 F 0.0
15:00 213.5 F 0.0
20:00 214.7 F 0.0
25:00 215.5 F 0.0
30:00 216.0 F 0.0
Drying Complete

FILLING: 1:23
HEATING: 20:35
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:59:13
CYCLE COMPLETE

03/27/2023 04:50 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____