

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3548

Name: RITTER 1  
Sterilizer ID: V2178198  
07/19/2023 10:25 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 177.7 F 0.0  
2:00 171.5 F 2.0  
4:00 181.6 F 4.9  
6:00 207.7 F 9.3  
8:00 223.0 F 7.4  
10:00 230.2 F 8.5  
12:00 240.4 F 13.5  
14:00 239.1 F 10.6  
16:00 237.1 F 8.3  
18:00 249.3 F 14.8  
20:00 260.5 F 21.1  
22:00 270.6 F 27.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.3  
1:00 272.5 F 29.0  
2:00 271.6 F 28.3  
3:00 272.1 F 28.9  
4:00 272.2 F 28.8

Min 271.2 F 28.2  
Max 272.5 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.0 F 0.0  
5:00 209.3 F 0.0  
10:00 230.9 F 0.0  
15:00 224.1 F 0.0  
20:00 221.1 F 0.0  
25:00 221.1 F 0.0  
30:00 219.7 F 0.0  
Drying Complete

FILLING: 1:05  
HEATING: 22:49  
STERILIZING: 4:00  
VENTING: 3:21  
DRYING: 30:00  
TOTAL CYCLE: 01:01:15  
CYCLE COMPLETE

07/19/2023 11:26 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_