

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3479

Name: RITTER2  
Sterilizer ID: V2313228  
03/27/2024 08:04 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

C232: FILL MODE  
WATER LOW  
CHAMBER TEMP 69F

\*\* ITEMS NOT STERILE \*\*

03/27/2024 08:09 AM