

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 002333  
12:03 31/08/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:34  
PRESSURIZING 3:35  
STERILIZING 4:42  
136.4 C 220kPa 4:42  
136.2 C 224kPa 5:13  
136.3 C 224kPa 5:43  
136.4 C 224kPa 6:13  
136.2 C 223kPa 6:43  
136.4 C 224kPa 7:13  
136.4 C 222kPa 7:43  
136.4 C 224kPa 8:13  
136.4 C 224kPa 8:13  
VENTING 8:13  
AIR DRYING 9:07  
CYCLE COMPLETE 9:12

-----  
Digital Signature #  
1E9895C5FB9F7ADD  
-----