

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2595

Name: RITTER2  
Sterilizer ID: V2313228  
06/30/2023 11:17 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 160.3 F 0.0  
2:00 162.3 F 0.9  
4:00 180.4 F 4.2  
6:00 209.7 F 8.7  
8:00 225.8 F 13.5  
10:00 226.8 F 7.6  
12:00 235.6 F 12.3  
14:00 235.1 F 9.4  
16:00 239.5 F 10.1  
18:00 247.2 F 14.2  
20:00 258.9 F 20.7  
22:00 269.7 F 27.6

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 28.9  
1:00 272.1 F 29.4  
2:00 272.5 F 29.4  
3:00 271.5 F 28.8  
4:00 272.5 F 29.8

Min 271.5 F 28.7  
Max 272.8 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.4 F 0.0  
5:00 204.9 F 0.0  
10:00 211.9 F 0.0  
15:00 214.2 F 0.0  
20:00 211.8 F 0.0  
25:00 212.1 F 0.0  
30:00 210.7 F 0.0  
Drying Complete

FILLING: 1:14  
HEATING: 22:52  
STERILIZING: 4:00  
VENTING: 3:27  
DRYING: 30:00  
TOTAL CYCLE: 01:01:34  
CYCLE COMPLETE

06/30/2023 12:18 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_