
Midmark M11 - v1.0.0
Total Cycles: 4443

Name: RITTER 1
Sterilizer ID: V2178198
04/10/2024 12:46 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 92.9 F 0.0
2:00 93.2 F 0.5
4:00 101.8 F 1.9
6:00 122.6 F 4.1
8:00 177.3 F 7.1
10:00 197.8 F 10.6
12:00 211.6 F 6.6
14:00 210.1 F 5.3
16:00 220.9 F 8.7
18:00 232.1 F 12.7
20:00 233.7 F 8.9
22:00 242.2 F 11.8
24:00 242.7 F 12.1
26:00 253.8 F 17.9
28:00 264.3 F 24.0
30:00 271.9 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.7
1:00 272.1 F 29.5
2:00 272.5 F 29.4
3:00 271.9 F 29.0
4:00 271.2 F 28.7

Min 271.0 F 28.6
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.9 F 0.0
5:00 208.0 F 0.0
10:00 205.0 F 0.0
15:00 195.8 F 0.0
20:00 193.3 F 0.0
25:00 194.6 F 0.0
30:00 216.5 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 30:13
STERILIZING: 4:00
VENTING: 3:47
DRYING: 30:00
TOTAL CYCLE: 01:08:56
CYCLE COMPLETE

04/10/2024 01:54 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____