

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2745

Name: RITTER2  
Sterilizer ID: V2313228  
08/09/2023 01:58 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 202.2 F 0.0  
2:00 190.8 F 4.3  
4:00 201.9 F 9.7  
6:00 220.3 F 6.6  
8:00 233.0 F 11.3  
10:00 234.0 F 9.0  
12:00 238.3 F 9.8  
14:00 248.1 F 14.9  
16:00 260.4 F 21.7  
18:00 271.4 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.1  
1:00 272.2 F 29.5  
2:00 271.8 F 29.6  
3:00 272.1 F 29.3  
4:00 272.0 F 29.7  
  
Min 271.4 F 29.0  
Max 272.7 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 204.6 F 0.0  
5:00 219.6 F 0.0  
10:00 225.0 F 0.0  
15:00 226.8 F 0.0  
20:00 221.2 F 0.0  
25:00 220.0 F 0.0  
30:00 216.1 F 0.0  
Drying Complete

FILLING: 2:00  
HEATING: 18:38  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 00:57:53  
CYCLE COMPLETE

08/09/2023 02:55 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_