

-----  
Brampton, ON  
Canada  
-----

STATIM 5000 S501R713  
SN 511019D00004  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 001452  
12:41 09/27/2021  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 5:03  
PRESSURIZING 6:52  
STERILIZING 8:13  
136.0 C 220kPa 8:13  
136.2 C 220kPa 8:43  
136.5 C 220kPa 9:13  
136.5 C 222kPa 9:43  
136.3 C 221kPa 10:13  
136.2 C 221kPa 10:43  
136.3 C 221kPa 11:13  
136.4 C 223kPa 11:43  
VENTING 11:43  
AIR DRYING 12:41  
CYCLE COMPLETE 42:41  
-----

Digital Signature #  
D7F515FC365CB2D4  
-----