

-----

Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.6uS / 0.3ppm  
CYCLE NUMBER 002686  
11:23 15/11/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:38  
PRESSURIZING 3:39  
STERILIZING 4:44  
136.4 C 220kPa 4:44  
136.2 C 223kPa 5:14  
136.3 C 223kPa 5:44  
136.2 C 222kPa 6:14  
136.4 C 223kPa 6:44  
136.4 C 224kPa 7:14  
136.4 C 221kPa 7:44  
136.2 C 223kPa 8:14  
136.2 C 223kPa 8:14  
VENTING 8:14  
AIR DRYING 9:09  
CYCLE COMPLETE 9:49

-----  
Digital Signature #  
709A1E8ADFEB93FB  
-----