
Midmark M11 - v1.0.0
Total Cycles: 5128

Name: RITTER 1
Sterilizer ID: V2178198
11/22/2024 02:10 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 112.9 F 0.0
2:00 120.3 F 0.0
4:00 136.4 F 0.0
6:00 182.7 F 0.2
8:00 202.8 F 0.6
10:00 215.2 F 3.7
12:00 224.7 F 7.3
14:00 232.3 F 11.2
16:00 231.2 F 7.2
18:00 243.0 F 12.7
20:00 241.0 F 11.0
22:00 238.4 F 9.8
24:00 251.9 F 16.3
26:00 263.2 F 22.8
28:00 272.7 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.3
1:00 272.3 F 28.9
2:00 272.7 F 29.0
3:00 272.1 F 28.6
4:00 271.6 F 28.2

Min 271.5 F 28.1
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.8 F 0.0
5:00 193.8 F 0.0
10:00 197.0 F 0.0
15:00 198.7 F 0.0
20:00 216.0 F 0.0
25:00 224.1 F 0.0
30:00 230.4 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 28:23
STERILIZING: 4:00
VENTING: 3:46
DRYING: 30:00
TOTAL CYCLE: 01:06:56
CYCLE COMPLETE

11/22/2024 03:16 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____