
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 005804
13:55 20/06/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:00
PRESSURIZING 3:01
STERILIZING 3:56
136.4 C 222kPa 3:56
136.2 C 224kPa 4:26
136.4 C 222kPa 4:56
136.2 C 221kPa 5:26
136.4 C 223kPa 5:56
136.4 C 225kPa 6:26
136.4 C 225kPa 6:56
136.4 C 225kPa 7:26
136.4 C 225kPa 7:26
VENTING 7:26
AIR DRYING 8:17
CYCLE COMPLETE 13:22

Digital Signature #
886E6523C78D6F23
