
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002478
11:01 06/02/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:04
PRESSURIZING 3:08
STERILIZING 4:10
136.4 C 223kPa 4:10
136.4 C 225kPa 4:40
136.4 C 225kPa 5:10
136.3 C 224kPa 5:40
136.4 C 222kPa 6:10
136.1 C 222kPa 6:40
136.3 C 221kPa 7:10
136.3 C 223kPa 7:40
136.3 C 223kPa 7:40
VENTING 7:40
AIR DRYING 8:31
CYCLE COMPLETE 10:10

Digital Signature #
613DFEC90D3D485C
