
Midmark M11 - v1.0.5
Total Cycles: 44

Name: RITTER2
Sterilizer ID: V2313228
05/26/2021 09:59 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 179.8 F 0.0
2:00 181.4 F 3.7
4:00 198.5 F 9.6
6:00 216.3 F 6.4
8:00 231.2 F 12.2
10:00 237.6 F 11.0
12:00 239.7 F 10.6
14:00 253.6 F 17.9
16:00 266.3 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.4
1:00 272.0 F 29.4
2:00 271.9 F 29.7
3:00 272.0 F 29.4
4:00 272.4 F 29.7

Min 271.2 F 29.3
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.8 F 0.0
5:00 237.6 F 0.0
10:00 226.2 F 0.0
15:00 231.2 F 0.0
20:00 225.7 F 0.0
25:00 230.8 F 0.0
30:00 224.2 F 0.0
Drying Complete

FILLING: 2:14
HEATING: 17:25
STERILIZING: 4:00
VENTING: 3:02
DRYING: 30:00
TOTAL CYCLE: 00:56:40
CYCLE COMPLETE

05/26/2021 10:55 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____