
Midmark M11 - v1.0.5
Total Cycles: 3416

Name: RITTER2
Sterilizer ID: V2313228
02/27/2024 03:01 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 92.1 F 0.0
2:00 93.0 F 0.0
4:00 100.0 F 0.5
6:00 132.6 F 2.4
8:00 158.2 F 5.0
10:00 176.2 F 8.0
12:00 206.6 F 11.8
14:00 215.8 F 5.1
16:00 221.6 F 9.0
18:00 233.0 F 13.2
20:00 235.2 F 10.1
22:00 235.8 F 9.1
24:00 246.1 F 14.1
26:00 256.9 F 19.9
28:00 267.0 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 29.3
1:00 272.2 F 30.3
2:00 272.1 F 29.8
3:00 272.7 F 30.3
4:00 271.9 F 30.2

Min 271.0 F 29.2
Max 273.3 F 30.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.1 F 0.0
5:00 213.8 F 0.0
10:00 219.3 F 0.0
15:00 218.6 F 0.0
20:00 221.8 F 0.0
25:00 227.1 F 0.0
30:00 229.5 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 29:38
STERILIZING: 4:00
VENTING: 3:38
DRYING: 30:00
TOTAL CYCLE: 01:08:10
CYCLE COMPLETE

02/27/2024 04:09 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____