
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 005572
9:26 23/05/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:36
PRESSURIZING 2:41
STERILIZING 3:42
136.0 C 220kPa 3:42
136.4 C 223kPa 4:12
136.2 C 222kPa 4:42
136.3 C 224kPa 5:12
136.4 C 225kPa 5:42
136.4 C 224kPa 6:12
136.4 C 220kPa 6:42
136.3 C 223kPa 7:12
136.3 C 223kPa 7:12
VENTING 7:12
AIR DRYING 8:05
CYCLE COMPLETE 8:13

Digital Signature #
9A5D826CE21FE988
