
Midmark M11 - v1.0.5
Total Cycles: 3437

Name: RITTER2
Sterilizer ID: V2313228
03/07/2024 09:13 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.3 F 0.0
2:00 75.2 F 0.0
4:00 98.9 F 0.0
6:00 107.3 F 1.1
8:00 143.0 F 3.6
10:00 168.1 F 6.5
12:00 184.6 F 10.5
14:00 203.7 F 5.4
16:00 215.3 F 6.0
18:00 225.2 F 9.8
20:00 233.1 F 10.2
22:00 237.4 F 10.7
24:00 235.0 F 8.2
26:00 248.1 F 14.6
28:00 259.1 F 20.8
30:00 269.3 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.8
1:00 272.4 F 29.4
2:00 271.5 F 28.8
3:00 272.5 F 29.5
4:00 271.6 F 28.7

Min 271.3 F 28.6
Max 272.6 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.3 F 0.0
5:00 230.1 F 0.0
10:00 211.0 F 0.0
15:00 209.4 F 0.0
20:00 197.8 F 0.0
25:00 199.7 F 0.0
30:00 208.9 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 31:05
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:09:11
CYCLE COMPLETE

03/07/2024 10:22 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____