
Midmark M11 - v1.0.5
Total Cycles: 3585

Name: RITTER2
Sterilizer ID: V2313228
05/02/2024 03:29 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 159.1 F 0.0
2:00 160.0 F 2.8
4:00 175.3 F 6.6
6:00 199.4 F 10.9
8:00 216.0 F 5.2
10:00 225.8 F 9.9
12:00 232.3 F 9.2
14:00 240.0 F 12.2
16:00 239.7 F 11.0
18:00 252.2 F 17.2
20:00 263.3 F 23.7
22:00 272.7 F 29.9

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.3
1:00 271.7 F 29.7
2:00 272.5 F 30.0
3:00 271.9 F 29.5
4:00 271.1 F 29.3

Min 271.1 F 29.2
Max 272.9 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.2 F 0.0
5:00 215.4 F 0.0
10:00 211.3 F 0.0
15:00 209.9 F 0.0
20:00 210.6 F 0.0
25:00 212.8 F 0.0
30:00 211.2 F 0.0
Drying Complete

FILLING: 1:46
HEATING: 22:11
STERILIZING: 4:00
VENTING: 3:37
DRYING: 30:00
TOTAL CYCLE: 01:01:34
CYCLE COMPLETE

05/02/2024 04:30 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____