
Midmark M11 - v1.0.0
Total Cycles: 958

Name: RITTER 1
Sterilizer ID: V2178198
07/05/2021 09:06 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 202.5 F 0.0
2:00 195.5 F 2.7
4:00 214.8 F 7.2
6:00 226.4 F 12.5
8:00 228.3 F 7.4
10:00 238.9 F 12.4
12:00 237.6 F 10.0
14:00 239.1 F 8.6
16:00 249.3 F 14.8
18:00 261.2 F 21.5
20:00 271.9 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.5 F 28.9
2:00 271.5 F 28.2
3:00 272.6 F 29.0
4:00 271.8 F 28.3

Min 271.4 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.1 F 0.0
5:00 232.8 F 0.0
10:00 228.0 F 0.0
15:00 228.9 F 0.0
20:00 228.6 F 0.0
25:00 227.2 F 0.0
30:00 192.4 F 0.0
Drying Complete

FILLING: 1:42
HEATING: 20:35
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:59:33
CYCLE COMPLETE

07/05/2021 10:05 AM
ACCEPTED: X
REJECTED:
Operator: U01