
Midmark M11 - v1.0.0
Total Cycles: 2291

Name: RITTER 1
Sterilizer ID: V2178198
09/06/2022 11:10 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 164.0 F 0.0
2:00 163.9 F 0.9
4:00 178.5 F 4.8
6:00 198.4 F 10.1
8:00 213.6 F 5.7
10:00 229.1 F 9.9
12:00 238.8 F 8.4
14:00 241.9 F 12.6
16:00 240.8 F 11.0
18:00 254.1 F 17.7
20:00 265.6 F 24.7

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.6
1:00 273.0 F 29.3
2:00 271.9 F 28.6
3:00 272.7 F 29.4
4:00 272.7 F 29.1

Min 271.7 F 28.5
Max 273.1 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.5 F 0.0
5:00 202.7 F 0.0
10:00 217.0 F 0.0
15:00 223.2 F 0.0
20:00 227.8 F 0.0
25:00 230.3 F 0.0
30:00 228.5 F 0.0
Drying Complete

FILLING: 1:33
HEATING: 21:45
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:00:33
CYCLE COMPLETE

09/06/2022 12:10 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____