
Midmark M11 - v1.0.0
Total Cycles: 4438

Name: RITTER 1
Sterilizer ID: V2178198
04/09/2024 10:01 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 128.6 F 0.0
2:00 127.3 F 0.9
4:00 129.4 F 3.1
6:00 151.1 F 6.0
8:00 184.9 F 9.2
10:00 196.9 F 13.5
12:00 210.1 F 5.1
14:00 219.1 F 8.8
16:00 230.4 F 12.9
18:00 233.0 F 8.8
20:00 242.9 F 12.5
22:00 242.2 F 11.7
24:00 253.3 F 17.3
26:00 263.6 F 23.3
28:00 272.9 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.5
1:00 272.3 F 29.4
2:00 272.5 F 29.2
3:00 272.0 F 28.8
4:00 271.4 F 28.5

Min 271.1 F 28.4
Max 272.8 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.8 F 0.0
5:00 215.3 F 0.0
10:00 216.7 F 0.0
15:00 214.0 F 0.0
20:00 216.7 F 0.0
25:00 223.9 F 0.0
30:00 232.4 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 28:25
STERILIZING: 4:00
VENTING: 3:44
DRYING: 30:00
TOTAL CYCLE: 01:07:08
CYCLE COMPLETE

04/09/2024 11:08 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____