

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5348

Name: RITTER 1  
Sterilizer ID: V2178198  
02/20/2025 11:47 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 176.7 F 0.0  
2:00 179.9 F 2.1  
4:00 185.7 F 5.8  
6:00 198.8 F 9.8  
8:00 219.9 F 7.2  
10:00 227.9 F 6.7  
12:00 233.9 F 11.1  
14:00 232.2 F 6.9  
16:00 244.3 F 12.7  
18:00 241.7 F 10.9  
20:00 253.4 F 16.9  
22:00 263.8 F 22.9  
24:00 272.7 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 28.4  
1:00 271.9 F 28.8  
2:00 272.4 F 29.0  
3:00 271.9 F 28.5  
4:00 271.5 F 28.7

Min 271.2 F 28.3  
Max 272.7 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 204.6 F 0.0  
5:00 231.1 F 0.0  
10:00 227.2 F 0.0  
15:00 205.0 F 0.0  
20:00 209.9 F 0.0  
25:00 216.2 F 0.0  
30:00 220.4 F 0.0  
Drying Complete

FILLING: 1:11  
HEATING: 25:01  
STERILIZING: 4:00  
VENTING: 3:57  
DRYING: 30:00  
TOTAL CYCLE: 01:04:09  
CYCLE COMPLETE

02/20/2025 12:51 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_