

-----  
Midmark M11 - v1.0.0  
Total Cycles: 770

Name: RITTER 1  
Sterilizer ID: V2178198  
04/06/2021 10:27 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 133.9 F 0.0  
2:00 135.1 F 0.5  
4:00 168.7 F 2.4  
6:00 189.0 F 6.5  
8:00 207.1 F 11.5  
10:00 223.8 F 5.6  
12:00 231.4 F 10.9  
14:00 231.5 F 7.2  
16:00 246.6 F 13.6  
18:00 243.8 F 11.8  
20:00 256.5 F 18.4  
22:00 267.7 F 25.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.5  
1:00 272.0 F 28.9  
2:00 272.3 F 28.8  
3:00 272.5 F 28.7  
4:00 271.9 F 28.5

Min 271.2 F 28.4  
Max 272.7 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 184.7 F 0.0  
5:00 231.8 F 0.0  
10:00 230.7 F 0.0  
15:00 232.3 F 0.0  
20:00 227.9 F 0.0  
25:00 231.8 F 0.0  
Dry Cycle Aborted

FILLING: 0:55  
HEATING: 23:52  
STERILIZING: 4:00  
VENTING: 3:00  
DRYING: 26:08  
TOTAL CYCLE: 00:57:54

04/06/2021 11:24 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01