
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001002
7:36 16/03/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:26
PRESSURIZING 2:30
STERILIZING 3:38
136.4 C 232kPa 3:38
136.4 C 232kPa 4:08
136.3 C 231kPa 4:38
136.4 C 230kPa 5:08
136.2 C 230kPa 5:38
136.2 C 232kPa 6:08
136.4 C 233kPa 6:38
136.1 C 230kPa 7:08
136.1 C 230kPa 7:08
VENTING 7:08
AIR DRYING 7:56
CYCLE COMPLETE 8:02

Digital Signature #
ED5B23C1CAE2F068
