
Midmark M11 - v1.0.5
Total Cycles: 2898

Name: RITTER2
Sterilizer ID: V2313228
09/18/2023 12:03 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 174.0 F 0.0
2:00 175.2 F 3.0
4:00 189.8 F 7.4
6:00 207.6 F 12.1
8:00 221.7 F 7.0
10:00 233.2 F 11.9
12:00 233.9 F 9.4
14:00 237.9 F 9.8
16:00 247.4 F 14.7
18:00 259.2 F 21.1
20:00 269.7 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.3
1:00 272.9 F 29.9
2:00 271.4 F 29.2
3:00 272.2 F 30.1
4:00 272.4 F 29.6

Min 271.2 F 29.2
Max 273.0 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.7 F 0.0
5:00 202.0 F 0.0
10:00 200.2 F 0.0
15:00 199.7 F 0.0
20:00 197.6 F 0.0
25:00 194.9 F 0.0
30:00 192.2 F 0.0
Drying Complete

FILLING: 1:46
HEATING: 20:59
STERILIZING: 4:00
VENTING: 3:41
DRYING: 30:00
TOTAL CYCLE: 01:00:26
CYCLE COMPLETE

09/18/2023 01:03 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____