
Midmark M11 - v1.0.0
Total Cycles: 5274

Name: RITTER 1
Sterilizer ID: V2178198
01/24/2025 11:07 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 150.1 F 0.0
2:00 150.0 F 2.1
4:00 167.8 F 5.5
6:00 188.8 F 9.3
8:00 207.7 F 11.2
10:00 220.7 F 6.6
12:00 232.4 F 11.2
14:00 230.5 F 6.7
16:00 243.6 F 12.4
18:00 240.5 F 10.3
20:00 252.6 F 16.4
22:00 263.3 F 22.7
24:00 272.4 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.2
1:00 272.4 F 28.9
2:00 271.6 F 28.2
3:00 271.8 F 28.8
4:00 271.8 F 28.6

Min 271.1 F 28.1
Max 272.5 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.7 F 0.0
5:00 214.5 F 0.0
10:00 197.3 F 0.0
15:00 199.4 F 0.0
20:00 212.5 F 0.0
25:00 215.9 F 0.0
30:00 217.3 F 0.0
Drying Complete

FILLING: 1:00
HEATING: 24:16
STERILIZING: 4:00
VENTING: 3:45
DRYING: 30:00
TOTAL CYCLE: 01:03:01
CYCLE COMPLETE

01/24/2025 12:09 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____