
Midmark M11 - v1.0.0
Total Cycles: 1908

Name: RITTER 1
Sterilizer ID: V2178198
05/04/2022 11:33 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 194.7 F 0.0
2:00 194.0 F 3.8
4:00 191.6 F 8.4
6:00 209.7 F 13.4
8:00 221.2 F 7.2
10:00 236.3 F 12.4
12:00 234.7 F 9.1
14:00 241.0 F 9.7
16:00 246.3 F 13.5
18:00 258.4 F 20.0
20:00 269.3 F 26.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.5
1:00 272.7 F 29.2
2:00 272.0 F 28.6
3:00 271.6 F 28.7
4:00 272.7 F 29.2

Min 271.4 F 28.3
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.2 F 0.0
5:00 230.4 F 0.0
10:00 221.9 F 0.0
15:00 228.3 F 0.0
20:00 233.0 F 0.0
25:00 231.2 F 0.0
30:00 227.7 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 21:07
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:59:21
CYCLE COMPLETE

05/04/2022 12:32 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____