

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2441

Name: RITTER 1  
Sterilizer ID: V2178198  
10/17/2022 12:09 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 190.8 F 0.0  
2:00 194.3 F 4.6  
4:00 204.2 F 9.5  
6:00 221.8 F 7.2  
8:00 229.4 F 8.8  
10:00 240.6 F 11.1  
12:00 241.0 F 11.6  
14:00 238.5 F 10.0  
16:00 252.2 F 16.8  
18:00 263.9 F 23.5  
20:00 271.5 F 28.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.6  
1:00 272.6 F 29.4  
2:00 272.1 F 28.8  
3:00 271.5 F 28.8  
4:00 272.6 F 29.4  
  
Min 271.4 F 28.5  
Max 272.6 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.2 F 0.0  
5:00 222.2 F 0.0  
10:00 233.3 F 0.0  
15:00 229.9 F 0.0  
20:00 225.9 F 0.0  
25:00 224.5 F 0.0  
30:00 223.6 F 0.0  
Drying Complete

FILLING: 1:40  
HEATING: 20:06  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 00:58:60  
CYCLE COMPLETE

10/17/2022 01:07 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_