

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 001427  
9:57 14/08/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:43  
PRESSURIZING 2:47  
STERILIZING 3:51  
136.4 C 220kPa 3:51  
136.2 C 224kPa 4:21  
136.4 C 226kPa 4:51  
136.4 C 223kPa 5:21  
136.1 C 224kPa 5:51  
136.3 C 224kPa 6:21  
136.2 C 225kPa 6:51  
136.2 C 223kPa 7:21  
136.2 C 223kPa 7:21  
VENTING 7:21  
AIR DRYING 8:11  
CYCLE COMPLETE 11:50

-----  
Digital Signature #  
DE8A44969A9B670F  
-----