
Midmark M11 - v1.0.5
Total Cycles: 1411

Name: RITTER2
Sterilizer ID: V2313228
08/24/2022 05:12 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 197.5 F 0.0
2:00 193.7 F 2.9
4:00 199.3 F 7.6
6:00 221.0 F 13.2
8:00 228.6 F 9.7
10:00 234.4 F 8.5
12:00 242.7 F 11.7
14:00 246.0 F 13.8
16:00 259.3 F 21.0
18:00 270.3 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.0
1:00 272.4 F 29.4
2:00 271.7 F 29.4
3:00 272.2 F 29.3
4:00 271.7 F 29.4

Min 271.4 F 28.9
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.5 F 0.0
5:00 217.5 F 0.0
10:00 223.1 F 0.0
15:00 228.4 F 0.0
20:00 227.6 F 0.0
25:00 227.1 F 0.0
30:00 225.1 F 0.0
Drying Complete

FILLING: 1:58
HEATING: 18:47
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 00:57:51
CYCLE COMPLETE

08/24/2022 06:09 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____