
Midmark M11 - v1.0.5
Total Cycles: 2579

Name: RITTER2
Sterilizer ID: V2313228
06/28/2023 10:05 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 170.0 F 0.0
2:00 171.1 F 0.1
4:00 187.0 F 3.6
6:00 213.5 F 9.0
8:00 223.7 F 7.8
10:00 234.9 F 11.3
12:00 235.0 F 9.3
14:00 237.4 F 9.2
16:00 249.8 F 15.6
18:00 262.5 F 22.9
20:00 272.1 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.9
1:00 272.6 F 29.7
2:00 271.9 F 29.0
3:00 271.6 F 29.2
4:00 272.7 F 29.7

Min 271.2 F 28.7
Max 272.8 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.8 F 0.0
5:00 212.5 F 0.0
10:00 224.4 F 0.0
15:00 227.5 F 0.0
20:00 227.8 F 0.0
25:00 229.7 F 0.0
30:00 228.5 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 20:12
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:58:15
CYCLE COMPLETE

06/28/2023 11:03 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____