
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000949
13:38 09/12/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:24
PRESSURIZING 2:27
STERILIZING 3:19
135.4 C 220kPa 3:19
136.2 C 222kPa 3:49
136.1 C 217kPa 4:19
136.3 C 219kPa 4:49
136.4 C 222kPa 5:19
136.3 C 222kPa 5:49
136.4 C 225kPa 6:19
136.4 C 225kPa 6:49
136.4 C 225kPa 6:49
VENTING 6:49
AIR DRYING 7:44
CYCLE COMPLETE 8:38

Digital Signature #
AC63AB061C4437AE
