
Midmark M11 - v1.0.0
Total Cycles: 2434

Name: RITTER 1
Sterilizer ID: V2178198
10/14/2022 01:19 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 174.1 F 0.0
2:00 171.2 F 4.1
4:00 191.4 F 8.3
6:00 215.1 F 13.6
8:00 224.2 F 7.1
10:00 233.3 F 12.2
12:00 235.2 F 9.1
14:00 239.7 F 9.5
16:00 246.8 F 13.7
18:00 258.7 F 20.2
20:00 269.8 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.7
1:00 272.6 F 29.4
2:00 271.7 F 28.7
3:00 272.5 F 29.2
4:00 271.8 F 28.8

Min 271.6 F 28.6
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.7 F 0.0
5:00 238.3 F 0.0
10:00 230.2 F 0.0
15:00 227.7 F 0.0
20:00 228.5 F 0.0
25:00 228.0 F 0.0
30:00 227.7 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 20:49
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:59:02
CYCLE COMPLETE

10/14/2022 02:17 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____