
Midmark M11 - v1.0.0
Total Cycles: 2765

Name: RITTER 1
Sterilizer ID: V2178198
01/11/2023 11:45 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 201.1 F 0.0
2:00 191.2 F 0.0
4:00 206.5 F 5.0
6:00 222.1 F 10.6
8:00 226.4 F 6.2
10:00 239.7 F 12.2
12:00 238.7 F 10.2
14:00 238.1 F 8.7
16:00 251.8 F 16.0
18:00 264.2 F 23.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.1
1:00 272.5 F 28.8
2:00 271.5 F 28.0
3:00 272.4 F 28.9
4:00 272.2 F 28.5

Min 271.3 F 27.9
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.4 F 0.0
5:00 222.3 F 0.0
10:00 221.8 F 0.0
15:00 224.5 F 0.0
20:00 228.7 F 0.0
25:00 233.2 F 0.0
30:00 216.7 F 0.0
Drying Complete

FILLING: 1:33
HEATING: 19:58
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:58:43
CYCLE COMPLETE

01/11/2023 12:43 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____