
Midmark M11 - v1.0.0
Total Cycles: 3415

Name: RITTER 1
Sterilizer ID: V2178198
06/20/2023 01:14 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 198.5 F 0.0
2:00 196.0 F 3.3
4:00 202.4 F 8.2
6:00 226.1 F 9.2
8:00 231.8 F 9.2
10:00 237.4 F 9.6
12:00 243.8 F 12.6
14:00 242.5 F 11.4
16:00 255.6 F 18.1
18:00 267.0 F 25.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.7 F 28.9
2:00 271.5 F 28.1
3:00 272.8 F 29.0
4:00 272.0 F 28.4

Min 271.4 F 28.0
Max 272.9 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.4 F 0.0
5:00 234.4 F 0.0
10:00 230.9 F 0.0
15:00 227.4 F 0.0
20:00 226.2 F 0.0
25:00 227.9 F 0.0
30:00 226.5 F 0.0

Drying Complete

FILLING: 1:52
HEATING: 19:29
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:58:31
CYCLE COMPLETE

06/20/2023 02:12 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____