
Midmark M11 - v1.0.5
Total Cycles: 730

Name: RITTER2
Sterilizer ID: V2313228
01/26/2022 10:52 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 153.4 F 0.0
2:00 164.2 F 3.0
4:00 189.3 F 7.6
6:00 211.1 F 12.9
8:00 224.5 F 7.1
10:00 238.1 F 12.7
12:00 239.3 F 11.0
14:00 239.3 F 10.2
16:00 253.3 F 17.2
18:00 265.1 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.7
1:00 272.7 F 29.4
2:00 271.8 F 28.6
3:00 272.2 F 29.4
4:00 272.6 F 29.2

Min 271.4 F 28.5
Max 272.8 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.5 F 0.0
5:00 205.9 F 0.0
10:00 212.5 F 0.0
15:00 214.7 F 0.0
20:00 211.2 F 0.0
25:00 213.0 F 0.0
30:00 213.9 F 0.0
Drying Complete

FILLING: 1:39
HEATING: 19:50
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:58:36
CYCLE COMPLETE

01/26/2022 11:50 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____