
Midmark M11 - v1.0.5
Total Cycles: 850

Name: RITTER2
Sterilizer ID: V2313228
03/01/2022 10:11 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 191.8 F 0.0
2:00 184.4 F 3.6
4:00 204.2 F 9.2
6:00 225.1 F 7.5
8:00 232.5 F 11.2
10:00 235.0 F 9.1
12:00 238.0 F 9.6
14:00 249.2 F 15.4
16:00 262.2 F 22.7
18:00 272.6 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.1
1:00 272.5 F 29.3
2:00 272.2 F 29.8
3:00 271.9 F 29.1
4:00 272.4 F 29.7

Min 271.3 F 28.9
Max 272.6 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.1 F 0.0
5:00 238.6 F 0.0
10:00 225.9 F 0.0
15:00 226.9 F 0.0
20:00 221.5 F 0.0
25:00 228.4 F 0.0
30:00 221.6 F 0.0
Drying Complete

FILLING: 2:10
HEATING: 18:13
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:57:32
CYCLE COMPLETE

03/01/2022 11:33 AM
ACCEPTED: X
REJECTED:
Operator: U02