
Midmark M11 - v1.0.0
Total Cycles: 4528

Name: RITTER 1
Sterilizer ID: V2178198
05/06/2024 02:30 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 126.8 F 0.0
2:00 128.7 F 1.5
4:00 146.9 F 4.3
6:00 172.8 F 7.4
8:00 195.7 F 11.4
10:00 210.1 F 3.5
12:00 220.6 F 7.1
14:00 230.4 F 11.1
16:00 230.1 F 6.8
18:00 241.7 F 12.3
20:00 239.9 F 10.4
22:00 251.9 F 16.4
24:00 262.5 F 22.5
26:00 272.6 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.7 F 29.1
2:00 271.8 F 28.3
3:00 272.1 F 28.9
4:00 272.4 F 28.9

Min 271.4 F 28.2
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.7 F 0.0
5:00 194.2 F 0.0
10:00 192.9 F 0.0
15:00 206.0 F 0.0
20:00 205.2 F 0.0
25:00 216.9 F 0.0
30:00 217.2 F 0.0
Drying Complete

FILLING: 1:42
HEATING: 26:26
STERILIZING: 4:00
VENTING: 3:45
DRYING: 30:00
TOTAL CYCLE: 01:05:53
CYCLE COMPLETE

05/06/2024 03:35 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____