

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R750  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000525  
7:05 29/09/2020

UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:14  
PRESSURIZING 3:18  
STERILIZING 4:32  
136.1 C 220kPa 4:32  
136.4 C 224kPa 5:02  
136.2 C 221kPa 5:32  
136.3 C 222kPa 6:02  
136.3 C 222kPa 6:32  
136.3 C 220kPa 7:02  
136.3 C 222kPa 7:32  
136.4 C 224kPa 8:02  
136.4 C 224kPa 8:02  
VENTING 8:02  
AIR DRYING 8:54  
CYCLE COMPLETE 8:56

-----  
Digital Signature #  
2145FAF8643151D9  
-----