

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4986

Name: RITTER 1  
Sterilizer ID: V2178198  
10/10/2024 12:27 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 146.5 F 0.0  
2:00 145.9 F 0.5  
4:00 170.7 F 1.8  
6:00 191.2 F 5.4  
8:00 205.5 F 8.9  
10:00 218.8 F 13.0  
12:00 226.0 F 7.7  
14:00 234.2 F 12.5  
16:00 236.7 F 9.8  
18:00 237.5 F 9.0  
20:00 248.4 F 14.4  
22:00 260.0 F 20.8  
24:00 270.4 F 27.6

STERILIZING  
mm:ss Degrees PSI  
0:00 272.2 F 28.8  
1:00 272.7 F 29.5  
2:00 272.1 F 28.8  
3:00 272.6 F 29.5  
4:00 272.4 F 29.0

Min 271.7 F 28.7  
Max 272.8 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.7 F 0.0  
5:00 222.5 F 0.0  
10:00 222.6 F 0.0  
15:00 232.3 F 0.0  
20:00 235.3 F 0.0  
25:00 228.4 F 0.0  
30:00 227.5 F 0.0  
Drying Complete

FILLING: 0:52  
HEATING: 25:33  
STERILIZING: 4:00  
VENTING: 3:55  
DRYING: 30:00  
TOTAL CYCLE: 01:04:20  
CYCLE COMPLETE

10/10/2024 01:31 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_