
Midmark M11 - v1.0.5
Total Cycles: 796

Name: RITTER2
Sterilizer ID: V2313228
02/14/2022 02:10 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 179.4 F 0.0
2:00 177.8 F 4.0
4:00 199.5 F 9.3
6:00 220.0 F 7.4
8:00 231.6 F 10.8
10:00 233.1 F 8.6
12:00 239.2 F 10.2
14:00 248.0 F 14.7
16:00 260.8 F 21.8
18:00 271.7 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.1
1:00 272.4 F 29.4
2:00 271.9 F 29.5
3:00 272.2 F 29.3
4:00 272.0 F 29.6

Min 271.4 F 28.9
Max 273.0 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.1 F 0.0
5:00 216.8 F 0.0
10:00 220.8 F 0.0
15:00 222.2 F 0.0
20:00 229.5 F 0.0
25:00 218.6 F 0.0
30:00 214.0 F 0.0
Drying Complete

FILLING: 2:05
HEATING: 18:33
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:57:45
CYCLE COMPLETE

02/14/2022 03:07 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____