
Midmark M11 - v1.0.5
Total Cycles: 1087

Name: RITTER2
Sterilizer ID: V2313228
05/10/2022 09:02 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 131.8 F 0.0
2:00 132.4 F 0.0
4:00 140.4 F 0.1
6:00 172.1 F 3.7
8:00 194.1 F 8.2
10:00 216.5 F 13.5
12:00 224.7 F 7.9
14:00 236.7 F 12.7
16:00 236.8 F 9.9
18:00 237.6 F 9.2
20:00 249.2 F 15.0
22:00 261.1 F 21.8
24:00 271.9 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.7
1:00 272.6 F 29.3
2:00 271.5 F 28.7
3:00 272.5 F 29.3
4:00 271.5 F 28.6

Min 271.4 F 28.6
Max 272.8 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.5 F 0.0
5:00 209.5 F 0.0
10:00 211.9 F 0.0
15:00 213.5 F 0.0
20:00 213.9 F 0.0
25:00 215.5 F 0.0
30:00 210.2 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 24:35
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:02:35
CYCLE COMPLETE

05/10/2022 10:04 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____