
Midmark M11 - v1.0.5
Total Cycles: 783

Name: RITTER2
Sterilizer ID: V2313228
02/10/2022 02:19 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 157.9 F 0.0
2:00 164.1 F 0.3
4:00 187.2 F 2.1
6:00 207.7 F 5.0
8:00 222.7 F 8.5
10:00 236.0 F 12.9
12:00 238.6 F 10.4
14:00 236.8 F 8.8
16:00 239.5 F 10.0
18:00 250.2 F 15.6
20:00 262.2 F 22.5
22:00 272.8 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.7
1:00 272.6 F 29.0
2:00 272.6 F 29.5
3:00 271.9 F 28.6
4:00 273.0 F 29.4

Min 271.6 F 28.5
Max 273.0 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.1 F 0.0
5:00 241.5 F 0.0
10:00 196.7 F 0.0
15:00 192.4 F 0.0
20:00 164.8 F 0.0
25:00 161.6 F 0.0
30:00 172.1 F 0.0
Drying Complete

FILLING: 1:39
HEATING: 22:18
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 01:01:01
CYCLE COMPLETE

02/10/2022 03:19 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____