
Midmark M11 - v1.0.5
Total Cycles: 16

Name: RITTER2
Sterilizer ID: V2313228
04/26/2021 09:55 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 149.6 F 0.0
2:00 152.4 F 2.2
4:00 169.8 F 6.1
6:00 193.4 F 11.1
8:00 214.7 F 5.1
10:00 229.3 F 10.4
12:00 233.2 F 8.2
14:00 241.8 F 11.3
16:00 247.9 F 14.5
18:00 260.5 F 21.5
20:00 272.0 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.6
1:00 272.4 F 29.0
2:00 272.0 F 29.1
3:00 272.0 F 28.7
4:00 272.2 F 29.3

Min 271.4 F 28.4
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.7 F 0.0
5:00 212.9 F 0.0
10:00 224.1 F 0.0
15:00 219.6 F 0.0
20:00 212.5 F 0.0
Dry Cycle Aborted

FILLING: 1:30
HEATING: 20:30
STERILIZING: 4:00
VENTING: 3:02
DRYING: 21:56
TOTAL CYCLE: 00:50:59

04/26/2021 10:46 AM
ACCEPTED: X
REJECTED:
Operator: U01