

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000931  
17:28 11/02/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:45  
PRESSURIZING 2:53  
STERILIZING 4:14  
136.3 C 222kPa 4:14  
136.4 C 233kPa 4:44  
136.2 C 231kPa 5:14  
136.2 C 232kPa 5:44  
136.4 C 232kPa 6:14  
136.4 C 233kPa 6:44  
136.4 C 230kPa 7:14  
136.4 C 233kPa 7:44  
136.4 C 233kPa 7:44  
VENTING 7:44  
AIR DRYING 8:40  
CYCLE COMPLETE 23:40

-----  
Digital Signature #  
10C33907E5F49638  
-----