

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1746

Name: RITTER 1  
Sterilizer ID: V2178198  
03/25/2022 10:33 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 154.4 F 0.0  
2:00 155.3 F 1.2  
4:00 170.3 F 4.1  
6:00 191.8 F 8.1  
8:00 210.3 F 13.0  
10:00 221.8 F 7.0  
12:00 234.1 F 12.0  
14:00 234.1 F 8.9  
16:00 239.4 F 10.3  
18:00 245.8 F 13.4  
20:00 258.0 F 19.9  
22:00 269.0 F 26.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.5  
1:00 272.4 F 29.0  
2:00 271.7 F 28.8  
3:00 272.3 F 28.9  
4:00 271.5 F 28.7

Min 271.4 F 28.4  
Max 272.6 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 197.4 F 0.0  
5:00 230.6 F 0.0  
10:00 230.0 F 0.0  
15:00 230.0 F 0.0  
20:00 229.7 F 0.0  
25:00 229.5 F 0.0  
30:00 223.9 F 0.0  
Drying Complete

FILLING: 1:03  
HEATING: 23:05  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 01:01:17  
CYCLE COMPLETE

03/25/2022 11:42 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02