
Midmark M11 - v1.0.5
Total Cycles: 405

Name: RITTER2
Sterilizer ID: V2313228
09/29/2021 08:37 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 99.5 F 0.0
2:00 101.2 F 0.6
4:00 125.5 F 2.3
6:00 185.3 F 4.7
8:00 159.7 F 8.6
10:00 180.6 F 12.8
12:00 206.6 F 6.4
14:00 221.3 F 10.4
16:00 233.0 F 8.4
18:00 237.0 F 11.2
20:00 235.4 F 8.8
22:00 247.1 F 14.6
24:00 257.5 F 20.3
26:00 267.3 F 26.6
28:00 273.2 F 30.4

STERILIZING
mm:ss Degrees PSI
0:00 272.2 F 30.1
1:00 272.5 F 30.5
2:00 272.3 F 30.3
3:00 272.4 F 30.2
4:00 272.3 F 30.0

Min 272.1 F 30.0
Max 272.8 F 30.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.6 F 0.0
5:00 240.1 F 0.0
10:00 226.7 F 0.0
15:00 217.7 F 0.0
20:00 210.7 F 0.0
25:00 213.4 F 0.0
30:00 196.6 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 28:13
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:06:28
CYCLE COMPLETE

09/29/2021 09:44 AM
ACCEPTED: X
REJECTED:
Operator: U01