
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000330
9:04 18/12/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:32
PRESSURIZING 2:42
STERILIZING 3:57
136.0 C 227kPa 3:57
136.0 C 230kPa 4:27
136.0 C 230kPa 4:57
136.3 C 231kPa 5:27
136.1 C 229kPa 5:57
136.1 C 229kPa 6:27
136.2 C 231kPa 6:57
136.4 C 232kPa 7:27
136.4 C 232kPa 7:27
VENTING 7:27
AIR DRYING 8:22
CYCLE COMPLETE 23:22

Digital Signature #
FDECC661A55E482C
